

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00624	Calib. Date:	09/09/2021
Certification File No.:	00625	Cert. Date:	09/09/2021
Linearity File No.:	00626	Lin. Date:	09/09/2021
Solution File No.:	00629	Soln. Date:	10/08/2021
Sequential File No.:	00629	File Date:	10/08/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0061
Solution Control Lot:	21010	Expires:	01/13/2023
		Bottle No.:	1272

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:45D		
Control 1 EC	-.---%	10:45D	22.0°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	10:45D	22.0°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

[Handwritten Signature] #421

Badge No.: 421

Date: 10/08/2021

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00624	Calib. Date:	09/09/2021
Certification File No.:	00625	Cert. No.:	00048
Linearity File No.:	00626	Cert. Date:	09/09/2021
Solution File No.:	00630	Lin. No.:	00037
Sequential File No.:	00630	Lin. Date:	09/09/2021
		Soln. Date:	10/08/2021
		Soln. No.:	00257
		File Date:	10/08/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0061
Solution Control Lot:	21010	Expires:	01/13/2023
		Bottle No.:	1272

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:14D		
Control 1 EC	0.101%	13:15D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:15D		
Control 2 EC	0.101%	13:16D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	13:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:17D		
Control 3 EC	0.101%	13:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	13:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

PFC. [Signature] #421

Badge No.: 421

Date: 10/08/2021