

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00584	Calib. Date: 07/14/2020
Certification File No.:	00585	Cert. No.: 00043
Linearity File No.:	00586	Cert. Date: 07/14/2020
Solution File No.:	00594	Lin. No.: 00034
Sequential File No.:	00594	Lin. Date: 07/14/2020
		Soln. No.: 00245
		Soln. Date: 12/03/2020
		File Date: 12/03/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0259

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:36S		
Control 1 EC	0.100%	07:37S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:38S		
Control 2 EC	0.098%	07:38S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:39S		
Control 3 EC	0.100%	07:40S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:40S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: PH1  #421

Badge No.: 421

Date: 12/03/2020