

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00606	Calib. Date: 03/31/2021
Certification File No.:	00607	Calib. No.: 00047
Linearity File No.:	00608	Cert. Date: 03/31/2021
Solution File No.:	00617	Cert. No.: 00036
Sequential File No.:	00617	Lin. Date: 03/31/2021
		Lin. No.: 00036
		Soln. Date: 05/26/2021
		Soln. No.: 00251
		File Date: 05/26/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	21010	Expires: 01/13/2023
		Bottle No.: 1450

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:19D		
Control 1 EC	0.100%	12:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:20D		
Control 2 EC	0.101%	12:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:22D		
Control 3 EC	0.101%	12:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 05/26/2021