

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	01106	Calib. Date: 01/25/2020
Certification File No.:	01107	Calib. No.: 00043
Linearity File No.:	01108	Cert. Date: 01/25/2020
Solution File No.:	01113	Cert. No.: 00031
Sequential File No.:	01113	Lin. Date: 01/25/2020
		Lin. No.: 00031
		Soln. Date: 02/14/2020
		Soln. No.: 00235
		File Date: 02/14/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	19020	Expires: 01/08/2021
		Bottle No.: 0281

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:58S		
Control 1 EC	0.097%	07:59S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	07:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:59S		
Control 2 EC	0.096%	08:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:01S		
Control 3 EC	0.098%	08:01S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 02/14/2020