

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	01008	Calib. Date:	01/14/2019
Certification File No.:	01009	Cert. Date:	01/14/2019
Linearity File No.:	01010	Lin. Date:	01/14/2019
Solution File No.:	01017	Soln. Date:	02/12/2019
Sequential File No.:	01017	File Date:	02/12/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0061
Solution Control Lot:	18150	Expires:	05/10/2020
		Bottle No.:	1410

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:58S		
Control 1 EC	0.102%	07:58S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	07:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:59S		
Control 2 EC	0.101%	08:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:00S		
Control 3 EC	0.101%	08:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 02/12/2019