

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	01106	Calib. Date:	01/25/2020
Certification File No.:	01107	Cert. No.:	00043
Linearity File No.:	01108	Cert. Date:	01/25/2020
Solution File No.:	01115	Lin. No.:	00031
Sequential File No.:	01115	Lin. Date:	01/25/2020
		Soln. Date:	03/09/2020
		Soln. No.:	00236
		File Date:	03/09/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0061
Solution Control Lot:	19020	Expires:	01/08/2021
		Bottle No.:	0330

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:16S		
Control 1 EC	0.099%	10:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:17S		
Control 2 EC	0.099%	10:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18S		
Control 3 EC	0.099%	10:19S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 03/09/2020