

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 01008 Calib. Date: 01/14/2019 Calib. No.: 00041
Certification File No.: 01009 Cert. Date: 01/14/2019 Cert. No.: 00029
Linearity File No.: 01010 Lin. Date: 01/14/2019 Lin. No.: 00029
Solution File No.: 01037 Soln. Date: 05/07/2019 Soln. No.: 00224
Sequential File No.: 01037 File Date: 05/07/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 05/10/2020
Solution Control Lot: 18150 Bottle No.: 0889

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:01D		
Control 1 EC	0.100%	12:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:02D		
Control 2 EC	0.100%	12:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:04D		
Control 3 EC	0.100%	12:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 05/07/2019