

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 01106 Calib. Date: 01/25/2020 Calib. No.: 00043  
Certification File No.: 01107 Cert. Date: 01/25/2020 Cert. No.: 00031  
Linearity File No.: 01108 Lin. Date: 01/25/2020 Lin. No.: 00031  
Solution File No.: 01120 Soln. Date: 06/04/2020 Soln. No.: 00239  
Sequential File No.: 01120 File Date: 06/04/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 02/19/2022  
Solution Control Lot: 20090 Bottle No.: 0516

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:04D		
Control 1 EC	0.101%	12:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:05D		
Control 2 EC	0.100%	12:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:06D		
Control 3 EC	0.100%	12:07D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: Sgt Gary A #351

Badge No.: 351

Date: 06/04/2020