

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	01008	Calib. Date: 01/14/2019
Certification File No.:	01009	Calib. No.: 00041
Linearity File No.:	01010	Cert. Date: 01/14/2019
Solution File No.:	01052	Cert. No.: 00029
Sequential File No.:	01052	Lin. Date: 01/14/2019
		Lin. No.: 00029
		Soln. Date: 07/03/2019
		Soln. No.: 00226
		File Date: 07/03/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 1350

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:10D		
Control 1 EC	0.101%	10:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:11D		
Control 2 EC	0.100%	10:12D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:12D		
Control 3 EC	0.100%	10:13D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

PH. [Signature] #421

Badge No.: 421

Date: 07/03/2019