

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
 Location: SOUTH BRUNSWICK POLICE
 Calibration File No.: 01053 Calib. Date: 07/10/2019 Calib. No.: 00042
 Certification File No.: 01054 Cert. Date: 07/10/2019 Cert. No.: 00030
 Linearity File No.: 01055 Lin. Date: 07/10/2019 Lin. No.: 00030
 Solution File No.: 01069 Soln. Date: 08/30/2019 Soln. No.: 00229
 Sequential File No.: 01069 File Date: 08/30/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
 Control Solution %: 0.100% Expires: 01/08/2021
 Solution Control Lot: 19020 Bottle No.: 0265

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 08:23D | | |
| Control 1 EC | 0.100% | 08:23D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 08:23D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:24D | | |
| Control 2 EC | 0.100% | 08:25D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 08:25D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:25D | | |
| Control 3 EC | 0.100% | 08:26D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 08:26D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:27D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Jesse Blake #421

Badge No.: 421

Date: 08/30/2019