

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	01053	Calib. Date: 07/10/2019
Certification File No.:	01054	Cert. No.: 00042
Linearity File No.:	01055	Cert. Date: 07/10/2019
Solution File No.:	01087	Lin. No.: 00030
Sequential File No.:	01087	Lin. Date: 07/10/2019
		Soln. No.: 00231
		Soln. Date: 10/24/2019
		File Date: 10/24/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	19020	Expires: 01/08/2021
		Bottle No.: 0341

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:37D		
Control 1 EC	0.100%	07:38D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:38D		
Control 2 EC	0.099%	07:39D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:39D		
Control 3 EC	0.100%	07:40D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_ 

Badge No.: 421

Date: 10/24/2019