

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 01053 Calib. Date: 07/10/2019 Calib. No.: 00042
Certification File No.: 01054 Cert. Date: 07/10/2019 Cert. No.: 00030
Linearity File No.: 01055 Lin. Date: 07/10/2019 Lin. No.: 00030
Solution File No.: 01095 Soln. Date: 11/21/2019 Soln. No.: 00232
Sequential File No.: 01095 File Date: 11/21/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
Control Solution %: 0.100% Expires: 01/08/2021
Solution Control Lot: 19020 Bottle No.: 0268

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:43S		
Control 1 EC	0.100%	07:43S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:44S		
Control 2 EC	0.100%	07:45S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:45S		
Control 3 EC	0.100%	07:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 11/21/2019