

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	01123	Calib. Date: 07/14/2020
Certification File No.:	01124	Calib. No.: 00044
Linearity File No.:	01125	Cert. Date: 07/14/2020
Solution File No.:	01130	Cert. No.: 00032
Sequential File No.:	01130	Lin. Date: 07/14/2020
		Lin. No.: 00032
		Soln. Date: 08/05/2020
		Soln. No.: 00242
		File Date: 08/05/2020

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0167

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:43D		
Control 1 EC	0.099%	04:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:44D		
Control 2 EC	0.099%	04:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:45D		
Control 3 EC	0.099%	04:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 08/05/2020