

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	Calib. No.:	00043
Calibration File No.:	01106	Cert. Date:	01/25/2020
Certification File No.:	01107	Lin. Date:	01/25/2020
Linearity File No.:	01108	Soln. Date:	04/06/2020
Solution File No.:	01117	File Date:	04/06/2020
Sequential File No.:	01117		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0061
Solution Control Lot:	20090	Expires:	02/19/2022
		Bottle No.:	0400

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:32D		
Control 1 EC	0.099%	11:33D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:33D		
Control 2 EC	0.099%	11:34D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35D		
Control 3 EC	0.100%	11:35D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: \_\_\_\_\_



Badge No.: 351

Date: 04/06/2020