

1974

1975

1976

1977

1978

1979

1980

1981

1982

1983

1984

1985

1986

1987

1988

1989

1990

1991

1992

1993

1994

1995

1996

1997

1998

1999

2000

2001

2002

2003

2004

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00570	Calib. Date: 01/25/2020
Certification File No.:	00571	Calib. No.: 00042
Linearity File No.:	00572	Cert. Date: 01/25/2020
Solution File No.:	00574	Cert. No.: 00033
Sequential File No.:	00574	Lin. Date: 01/25/2020
		Lin. No.: 00033
		Soln. Date: 02/21/2020
		Soln. No.: 00232
		File Date: 02/21/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	19020	Expires: 01/08/2021
		Bottle No.: 0267

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:06S		
Control 1 EC	0.095%	08:07S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:08S		
Control 2 EC	0.095%	08:09S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:10S		
Control 3 EC	0.095%	08:11S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:12S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 02/21/2020