

Calibrating Unit New Standard Solution Report

| | | |
|-------------------------|------------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARTL-0014 |
| Location: | SOUTH BRUNSWICK POLICE | |
| Calibration File No.: | 00570 | Calib. Date: 01/25/2020 |
| Certification File No.: | 00571 | Calib. No.: 00042 |
| Linearity File No.: | 00572 | Cert. Date: 01/25/2020 |
| Solution File No.: | 00576 | Cert. No.: 00033 |
| Sequential File No.: | 00576 | Lin. Date: 01/25/2020 |
| | | Lin. No.: 00033 |
| | | Soln. Date: 03/09/2020 |
| | | Soln. No.: 00233 |
| | | File Date: 03/09/2020 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUH S3-0079 |
| Solution Control Lot: | 19020 | Expires: 01/08/2021 |
| | | Bottle No.: 0324 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 10:15S | | |
| Control 1 EC | 0.097% | 10:15S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 10:15S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:17S | | |
| Control 2 EC | 0.095% | 10:17S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 10:17S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:18S | | |
| Control 3 EC | 0.096% | 10:19S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 10:19S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:20S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 03/09/2020