

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
**Location:** SOUTH BRUNSWICK POLICE  
**Calibration File No.:** 00546 **Calib. Date:** 01/14/2019 **Calib. No.:** 00040  
**Certification File No.:** 00547 **Cert. Date:** 01/14/2019 **Cert. No.:** 00031  
**Linearity File No.:** 00548 **Lin. Date:** 01/14/2019 **Lin. No.:** 00031  
**Solution File No.:** 00551 **Soln. Date:** 03/05/2019 **Soln. No.:** 00217  
**Sequential File No.:** 00551 **File Date:** 03/05/2019

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDUH S3-0079  
**Control Solution %:** 0.100% **Expires:** 05/10/2020  
**Solution Control Lot:** 18150 **Bottle No.:** 1465

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:21S		
Control 1 EC	0.100%	12:21S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	12:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:22S		
Control 2 EC	0.098%	12:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:24S		
Control 3 EC	0.099%	12:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	12:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: SGT. Gary AH 351

Badge No.: 351

Date: 03/05/2019