

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00641 Calib. Date: 02/10/2022 Calib. No.: 00049
Certification File No.: 00642 Cert. Date: 02/10/2022 Cert. No.: 00038
Linearity File No.: 00643 Lin. Date: 02/10/2022 Lin. No.: 00038
Solution File No.: 00646 Soln. Date: 03/11/2022 Soln. No.: 00263
Sequential File No.: 00646 File Date: 03/11/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
Control Solution %: 0.100% Expires: 09/01/2023
Solution Control Lot: 21360 Bottle No.: 1437

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:47S		
Control 1 EC	0.098%	07:47S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:48S		
Control 2 EC	0.098%	07:48S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:48S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:49S		
Control 3 EC	0.098%	07:50S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

DFC. JBL #421

Badge No.: 421

Date: 03/11/2022