

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00570	Calib. Date: 01/25/2020
Certification File No.:	00571	Cert. No.: 00042
Linearity File No.:	00572	Cert. Date: 01/25/2020
Solution File No.:	00577	Lin. Date: 01/25/2020
Sequential File No.:	00577	Lin. No.: 00033
		Soln. Date: 04/06/2020
		Soln. No.: 00234
		File Date: 04/06/2020

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUH S3-0079
Control Solution %:	0.100%		Expires: 02/19/2022
Solution Control Lot:	20090		Bottle No.: 0056

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:33D		
Control 1 EC	0.102%	13:33D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:34D		
Control 2 EC	0.098%	13:35D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:36D		
Control 3 EC	0.099%	13:37D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 04/06/2020