

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00570	Calib. Date: 01/25/2020
Certification File No.:	00571	Calib. No.: 00042
Linearity File No.:	00572	Cert. Date: 01/25/2020
Solution File No.:	00580	Cert. No.: 00033
Sequential File No.:	00580	Lin. Date: 01/25/2020
		Lin. No.: 00033
		Soln. Date: 06/04/2020
		Soln. No.: 00237
		File Date: 06/04/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	20090	Expires: 02/19/2022
		Bottle No.: 0283

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:18D		
Control 1 EC	0.105%	13:19D	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.100%	13:19D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	13:20D		
Control 2 EC	0.103%	13:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:21D		
Control 3 EC	0.102%	13:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:23D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Sgt. Gary A #351

Badge No.: 351

Date: 06/04/2020

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00570	Calib. Date: 01/25/2020
Certification File No.:	00571	Calib. No.: 00042
Linearity File No.:	00572	Cert. Date: 01/25/2020
Solution File No.:	00579	Cert. No.: 00033
Sequential File No.:	00579	Lin. Date: 01/25/2020
		Lin. No.: 00033
		Soln. Date: 06/04/2020
		Soln. No.: 00236
		File Date: 06/04/2020

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUH S3-0079
Control Solution %: 0.100%		Expires: 02/19/2022
Solution Control Lot: 20090		Bottle No.: 0573

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:32D		
Control 1 EC	-.---%	11:33D	22.8°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	11:33D	22.8°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 06/04/2020

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00570	Calib. Date:	01/25/2020
Certification File No.:	00571	Cert. Date:	01/25/2020
Linearity File No.:	00572	Lin. Date:	01/25/2020
Solution File No.:	00581	Soln. Date:	06/04/2020
Sequential File No.:	00581	File Date:	06/04/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0079
Solution Control Lot:	19020	Expires:	01/08/2021
		Bottle No.:	0282

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:45D		
Control 1 EC	0.103%	15:45D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:46D		
Control 2 EC	0.100%	15:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:48D		
Control 3 EC	0.100%	15:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Sgt. Gary A. #351

Badge No.: 351

Date: 06/04/2020