

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00584 Calib. Date: 07/14/2020 Calib. No.: 00043  
Certification File No.: 00585 Cert. Date: 07/14/2020 Cert. No.: 00034  
Linearity File No.: 00586 Lin. Date: 07/14/2020 Lin. No.: 00034  
Solution File No.: 00593 Soln. Date: 11/06/2020 Soln. No.: 00244  
Sequential File No.: 00593 File Date: 11/06/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080  
Control Solution %: 0.100% Expires: 02/19/2022  
Solution Control Lot: 20090 Bottle No.: 0029

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:37S		
Control 1 EC	0.101%	07:37S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:38S		
Control 2 EC	0.099%	07:38S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:39S		
Control 3 EC	0.100%	07:40S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 11/06/2020