

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00560	Calib. Date: 07/10/2019
Certification File No.:	00561	Calib. No.: 00041
Linearity File No.:	00562	Cert. Date: 07/10/2019
Solution File No.:	00569	Cert. No.: 00032
Sequential File No.:	00569	Lin. Date: 07/10/2019
		Lin. No.: 00032
		Soln. Date: 12/18/2019
		Soln. No.: 00230
		File Date: 12/18/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	19020	Expires: 01/08/2021
		Bottle No.: 0338

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:20S		
Control 1 EC	0.098%	07:21S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:22S		
Control 2 EC	0.096%	07:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:24S		
Control 3 EC	0.097%	07:24S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 12/18/2019