

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00641	Calib. Date: 02/10/2022
Certification File No.:	00642	Calib. No.: 00049
Linearity File No.:	00643	Cert. Date: 02/10/2022
Solution File No.:	00656	Cert. No.: 00038
Sequential File No.:	00656	Lin. Date: 02/10/2022
		Lin. No.: 00038
		Soln. Date: 05/06/2022
		Soln. No.: 00265
		File Date: 05/06/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	21360	Expires: 09/01/2023
		Bottle No.: 1098

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:10D		
Control 1 EC	0.099%	08:11D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:11D		
Control 2 EC	0.098%	08:12D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:13D		
Control 3 EC	0.098%	08:13D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 05/06/2022