

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00570	Calib. Date:	01/25/2020
Certification File No.:	00571	Cert. Date:	01/25/2020
Linearity File No.:	00572	Lin. Date:	01/25/2020
Solution File No.:	00582	Soln. Date:	06/30/2020
Sequential File No.:	00582	File Date:	06/30/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0079
Solution Control Lot:	20090	Expires:	02/19/2022
		Bottle No.:	0230

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:27D		
Control 1 EC	0.104%	17:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:29D		
Control 2 EC	0.102%	17:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:30D		
Control 3 EC	0.102%	17:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: SFC Gary H #351

Badge No.: 351

Date: 06/30/2020