

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00584	Calib. Date: 07/14/2020
Certification File No.:	00585	Cert. No.: 00043
Linearity File No.:	00586	Cert. Date: 07/14/2020
Solution File No.:	00595	Lin. No.: 00034
Sequential File No.:	00595	Soln. No.: 00246
		File Date: 12/18/2020

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUH S3-0080
Control Solution %:	0.100%		Expires: 02/19/2022
Solution Control Lot:	20090		Bottle No.: 0941

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:26S		
Control 1 EC	0.098%	07:26S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:28S		
Control 2 EC	0.096%	07:28S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:29S		
Control 3 EC	0.097%	07:30S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 12/18/2020