

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00624 Calib. Date: 09/09/2021 Calib. No.: 00048  
Certification File No.: 00625 Cert. Date: 09/09/2021 Cert. No.: 00037  
Linearity File No.: 00626 Lin. Date: 09/09/2021 Lin. No.: 00037  
Solution File No.: 00631 Soln. Date: 11/05/2021 Soln. No.: 00258  
Sequential File No.: 00631 File Date: 11/05/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 01/13/2023  
Solution Control Lot: 21010 Bottle No.: 1284

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:26S		
Control 1 EC	0.101%	15:26S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.103%	15:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:27S		
Control 2 EC	0.101%	15:27S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	15:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:28S		
Control 3 EC	0.101%	15:29S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	15:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 11/05/2021