

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00524	Calib. Date:	03/27/2018
Certification File No.:	00525	Cert. Date:	03/27/2018
Linearity File No.:	00526	Lin. Date:	03/27/2018
Solution File No.:	00530	Soln. Date:	06/27/2018
Sequential File No.:	00530	File Date:	06/27/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0079
Solution Control Lot:	17420	Expires:	12/11/2019
		Bottle No.:	0980

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:32D		
Control 1 EC	0.100%	07:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:33D		
Control 2 EC	0.099%	07:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:34D		
Control 3 EC	0.099%	07:35D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: 

Badge No.: 421

Date: 06/27/2018