

Calibrating Unit


New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00624	Calib. Date: 09/09/2021
Certification File No.:	00625	Calib. No.: 00048
Linearity File No.:	00626	Cert. Date: 09/09/2021
Solution File No.:	00635	Cert. No.: 00037
Sequential File No.:	00635	Lin. Date: 09/09/2021
		Lin. No.: 00037
		Soln. Date: 12/29/2021
		Soln. No.: 00260
		File Date: 12/29/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	21010	Expires: 01/13/2023
		Bottle No.: 1433

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:40S		
Control 1 EC	0.103%	09:40S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41S		
Control 2 EC	0.102%	09:42S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.103%	09:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:42S		
Control 3 EC	0.102%	09:43S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	09:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:			
Last Name: MARRERO	First Name: JOSEPH		MI: A
Signature: 		Badge No.: 406	
		Date: 12/29/2021	