

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00624 **Calib. Date:** 09/09/2021 **Calib. No.:** 00048
Certification File No.: 00625 **Cert. Date:** 09/09/2021 **Cert. No.:** 00037
Linearity File No.: 00626 **Lin. Date:** 09/09/2021 **Lin. No.:** 00037
Solution File No.: 00634 **Soln. Date:** 12/03/2021 **Soln. No.:** 00259
Sequential File No.: 00634 **File Date:** 12/03/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0061
Control Solution %: 0.100% **Expires:** 01/13/2023
Solution Control Lot: 21010 **Bottle No.:** 1199

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:07S		
Control 1 EC	0.100%	08:08S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:08S		
Control 2 EC	0.101%	08:09S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:10S		
Control 3 EC	0.101%	08:10S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	08:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 12/03/2021