

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00624	Calib. Date:	09/09/2021
Certification File No.:	00625	Cert. Date:	09/09/2021
Linearity File No.:	00626	Lin. Date:	09/09/2021
Solution File No.:	00639	Soln. Date:	01/26/2022
Sequential File No.:	00639	File Date:	01/26/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0061
Solution Control Lot:	21010	Expires:	01/13/2023
		Bottle No.:	1430

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:04S		
Control 1 EC	0.102%	11:05S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:05S		
Control 2 EC	0.102%	11:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:07S		
Control 3 EC	0.102%	11:07S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.103%	11:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARRERO

First Name: JOSEPH

MI: A

Signature: _____

Badge No.: 406

Date: 01/26/2022