

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00531	Calib. Date:	10/22/2012
Certification File No.:	00532	Cert. No.:	00025
Linearity File No.:	00533	Cert. Date:	10/22/2012
Solution File No.:	00534	Lin. Date:	10/22/2012
Sequential File No.:	00534	Lin. No.:	00015
		Soln. Date:	10/22/2012
		Soln. No.:	00125
		File Date:	10/22/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0061
Solution Control Lot:	11I095	Expires:	09/15/2013
		Bottle No.:	0431

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:41D		
Control 1 EC	0.100%	14:42D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:42D		
Control 2 EC	0.100%	14:43D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:44D		
Control 3 EC	0.100%	14:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number DDUJP2-215 (MA)

Changed By:

Last Name: GONCALVES

First Name: MICHELLE

MI: L

Signature: Det I Silva # 6040

Badge No.: 6040

Date: 10/22/2012