

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00584	Calib. Date: 07/14/2020
Certification File No.:	00585	Calib. No.: 00043
Linearity File No.:	00586	Cert. Date: 07/14/2020
Solution File No.:	00587	Cert. No.: 00034
Sequential File No.:	00587	Lin. Date: 07/14/2020
		Lin. No.: 00034
		Soln. No.: 00240
		File Date: 07/14/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	19020	Expires: 01/08/2021
		Bottle No.: 0347

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:42D		
Control 1 EC	0.102%	11:43D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	11:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43D		
Control 2 EC	0.101%	11:44D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	11:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45D		
Control 3 EC	0.102%	11:45D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: _____

DDEE P2-022

MRW

Changed By:

Last Name: WATSON

First Name: MATTHEW

MI: R

Signature: _____

Tpr. I. Matthew W. Watson #7078

Badge No.: 7078

Date: 07/14/2020