

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00893 **Calib. Date:** 01/04/2017 **Calib. No.:** 00034
Certification File No.: 00894 **Cert. Date:** 01/04/2017 **Cert. No.:** 00024
Linearity File No.: 00895 **Lin. Date:** 01/04/2017 **Lin. No.:** 00024
Solution File No.: 00903 **Soln. Date:** 03/02/2017 **Soln. No.:** 00191
Sequential File No.: 00903 **File Date:** 03/02/2017

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0060
Control Solution %: 0.100% **Expires:** 07/06/2018
Solution Control Lot: 16150 **Bottle No.:** 1135

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:23S		
Control 1 EC	0.097%	07:23S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:24S		
Control 2 EC	0.098%	07:25S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:26S		
Control 3 EC	0.099%	07:27S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 03/02/2017