

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	01210	Calib. Date:	02/10/2022
Certification File No.:	01211	Cert. No.:	00049
Linearity File No.:	01212	Cert. Date:	02/10/2022
Solution File No.:	01246	Lin. Date:	02/10/2022
Sequential File No.:	01246	Soln. Date:	06/24/2022
		File Date:	06/24/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0080
Solution Control Lot:	21360	Expires:	09/01/2023
		Bottle No.:	1223

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:48D		
Control 1 EC	0.099%	15:48D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	15:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:49D		
Control 2 EC	0.097%	15:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:50D		
Control 3 EC	0.097%	15:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 06/24/2022