

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 01250 Calib. Date: 06/28/2022 Calib. No.: 00050
Certification File No.: 01251 Cert. Date: 06/28/2022 Cert. No.: 00037
Linearity File No.: 01252 Lin. Date: 06/28/2022 Lin. No.: 00037
Solution File No.: 01263 Soln. Date: 08/18/2022 Soln. No.: 00271
Sequential File No.: 01263 File Date: 08/18/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079
Control Solution %: 0.100% Expires: 09/01/2023
Solution Control Lot: 21360 Bottle No.: 0906

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:34D		
Control 1 EC	0.098%	11:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35D		
Control 2 EC	0.099%	11:36D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:37D		
Control 3 EC	0.098%	11:37D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: _____

FFC / Jh 408

Badge No.: 408

Date: 08/18/2022