

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	01282	Calib. Date:	12/12/2022
Certification File No.:	01283	Cert. No.:	00051
Linearity File No.:	01284	Cert. Date:	12/12/2022
Solution File No.:	01287	Lin. No.:	00038
Sequential File No.:	01287	Lin. Date:	12/12/2022
		Soln. Date:	01/05/2023
		Soln. No.:	00278
		File Date:	01/05/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0080
Solution Control Lot:	21360	Expires:	09/01/2023
		Bottle No.:	1102

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:28S		
Control 1 EC	0.100%	08:29S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:30S		
Control 2 EC	0.100%	08:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:31S		
Control 3 EC	0.100%	08:31S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: _____

PFC JG 408

Badge No.: 408

Date: 01/05/2023