

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	01250	Calib. Date:	06/28/2022
Certification File No.:	01251	Cert. Date:	06/28/2022
Linearity File No.:	01252	Lin. Date:	06/28/2022
Solution File No.:	01279	Soln. Date:	12/06/2022
Sequential File No.:	01279	File Date:	12/06/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0079
Solution Control Lot:	21360	Expires:	09/01/2023
		Bottle No.:	1229

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:11S		
Control 1 EC	----%	08:11S	34.0°C	CTRL GAS SUPPLY
Control 1 IR	----%	08:11S	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	----%	--:--		
Control 2 EC	----%	--:--	N/A	
Control 2 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		
Control 3 EC	----%	--:--	N/A	
Control 3 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: _____

PFC J M 408

Badge No.: 408

Date: 12/06/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	01250	Calib. Date: 06/28/2022
Certification File No.:	01251	Calib. No.: 00050
Linearity File No.:	01252	Cert. Date: 06/28/2022
Solution File No.:	01280	Cert. No.: 00037
Sequential File No.:	01280	Lin. Date: 06/28/2022
		Lin. No.: 00037
		Soln. Date: 12/06/2022
		Soln. No.: 00276
		File Date: 12/06/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	21360	Expires: 09/01/2023
		Bottle No.: 1099

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:08S		
Control 1 EC	0.099%	11:09S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:10S		
Control 2 EC	0.099%	11:10S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:11S		
Control 3 EC	0.100%	11:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:12S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMN

First Name: JASON

MI:

Signature: _____

PFC PFC J M 408

Badge No.: 408

Date: 12/06/2022