

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	01282	Calib. Date:	12/12/2022
Certification File No.:	01283	Cert. No.:	00051
Linearity File No.:	01284	Lin. Date:	12/12/2022
Solution File No.:	01294	Cert. No.:	00038
Sequential File No.:	01294	Lin. No.:	00038
		Soln. Date:	03/02/2023
		Soln. No.:	00280
		File Date:	03/02/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0080
Solution Control Lot:	23010	Expires:	01/11/2025
		Bottle No.:	0931

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:30S		
Control 1 EC	-.---%	06:31S	21.6°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	06:31S	21.6°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GASSMAN

First Name: JASON

MI:

Signature: \_\_\_\_\_

*PFC J M 408*

Badge No.: 408

Date: 03/02/2023

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 01282 Calib. Date: 12/12/2022 Calib. No.: 00051  
Certification File No.: 01283 Cert. Date: 12/12/2022 Cert. No.: 00038  
Linearity File No.: 01284 Lin. Date: 12/12/2022 Lin. No.: 00038  
Solution File No.: 01295 Soln. Date: 03/02/2023 Soln. No.: 00281  
Sequential File No.: 01295 File Date: 03/02/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080  
Control Solution %: 0.100% Expires: 01/11/2025  
Solution Control Lot: 23010 Bottle No.: 0892

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:00S		
Control 1 EC	0.101%	08:01S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:01S		
Control 2 EC	0.100%	08:02S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:02S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:03S		
Control 3 EC	0.100%	08:03S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GASSMAN

First Name: JASON

MI:

Signature: \_\_\_\_\_

*PFC J M 408*

Badge No.: 408

Date: 03/02/2023