

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 01282 Calib. Date: 12/12/2022 Calib. No.: 00051  
Certification File No.: 01283 Cert. Date: 12/12/2022 Cert. No.: 00038  
Linearity File No.: 01284 Lin. Date: 12/12/2022 Lin. No.: 00038  
Solution File No.: 01301 Soln. Date: 03/31/2023 Soln. No.: 00282  
Sequential File No.: 01301 File Date: 03/31/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080  
Control Solution %: 0.100% Expires: 01/11/2025  
Solution Control Lot: 23010 Bottle No.: 1025

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:55S		
Control 1 EC	0.100%	10:56S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:57S		
Control 2 EC	0.100%	10:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:58S		
Control 3 EC	0.100%	10:59S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 03/31/2023