Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C

Location: SOUTH BRUNSWICK POLICE

Serial No.: ARNK-0074

Calibration File No.:	01303	Calib. Date:	04/11/2023	Calib. No.:	00052
Certification File No.:	01304	Cert. Date:	04/11/2023	Cert. No.:	00039
Linearity File No.:	01305	Lin. Date:	04/11/2023	Lin. No.:	00039
Solution File No.:	01308	Soln. Date:	05/04/2023	Soln. No.:	00284
Sequential File No.:	01308	File Date:	05/04/2023		

Calibrating Unit: WET

Control Solution %: 0.100%

Solution Control Lot: 23010

Model No.: CU-34

Serial No.: DDUH S3-0079

Expires: 01/11/2025

Bottle No.: 1015

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:07D		
Control 1 EC	0.101%	13:07D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:08D		
Control 2 EC	0.100%	13:09D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:10D		
Control 3 EC	0.100%	13:10D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE First Name: JESSE MI: R

Signature: PFC. #421

Badge No.: 421

Date: 05/04/2023