

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00684 Calib. Date: 12/12/2022 Calib. No.: 00051
Certification File No.: 00685 Cert. Date: 12/12/2022 Cert. No.: 00040
Linearity File No.: 00686 Lin. Date: 12/12/2022 Lin. No.: 00040
Solution File No.: 00700 Soln. Date: 04/05/2023 Soln. No.: 00280
Sequential File No.: 00700 File Date: 04/05/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
Control Solution %: 0.100% Expires: 01/11/2025
Solution Control Lot: 23010 Bottle No.: 0947

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:24D		
Control 1 EC	0.099%	07:25D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:25D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:26D		
Control 2 EC	0.099%	07:26D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:26D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:27D		
Control 3 EC	0.099%	07:27D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:27D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: PFC. JBL #421

Badge No.: 421

Date: 04/05/2023