

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00701	Calib. Date: 04/11/2023
Certification File No.:	00702	Calib. No.: 00052
Linearity File No.:	00703	Cert. Date: 04/11/2023
Solution File No.:	00707	Cert. No.: 00041
Sequential File No.:	00707	Lin. Date: 04/11/2023
		Lin. No.: 00041
		Soln. Date: 05/10/2023
		Soln. No.: 00282
		File Date: 05/10/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	23010	Expires: 01/11/2025
		Bottle No.: 0883

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:49D		
Control 1 EC	0.102%	10:49D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:50D		
Control 2 EC	0.101%	10:51D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:51D		
Control 3 EC	0.101%	10:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GASSMAN

First Name: JASON

MI:

Signature: \_\_\_\_\_

*PFC J. Man 408*

Badge No.: 408

Date: 05/10/2023