

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00714 Calib. Date: 06/28/2023 Calib. No.: 00053
Certification File No.: 00715 Cert. Date: 06/28/2023 Cert. No.: 00042
Linearity File No.: 00716 Lin. Date: 06/28/2023 Lin. No.: 00042
Solution File No.: 00724 Soln. Date: 08/23/2023 Soln. No.: 00286
Sequential File No.: 00724 File Date: 08/23/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 01/11/2025
Solution Control Lot: 23010 Bottle No.: 0886

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:23D		
Control 1 EC	0.100%	07:23D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:24D		
Control 2 EC	0.100%	07:25D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:25D		
Control 3 EC	0.101%	07:26D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: _____

PFC J M 408

Badge No.: 408

Date: 08/23/2023