

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00714	Calib. Date: 06/28/2023
Certification File No.:	00715	Calib. No.: 00053
Linearity File No.:	00716	Cert. Date: 06/28/2023
Solution File No.:	00722	Cert. No.: 00042
Sequential File No.:	00722	Lin. Date: 06/28/2023
		Lin. No.: 00042
		Soln. Date: 07/26/2023
		Soln. No.: 00285
		File Date: 07/26/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	23010	Expires: 01/11/2025
		Bottle No.: 0728

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:56D		
Control 1 EC	0.099%	08:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:58D		
Control 2 EC	0.099%	08:58D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:59D		
Control 3 EC	0.101%	08:59D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

PFC. [Signature] #421

Badge No.: 421

Date: 07/26/2023