

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|------------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARNK-0074 |
| Location: | SOUTH BRUNSWICK POLICE | | |
| Calibration File No.: | 01316 | Calib. Date: | 06/28/2023 |
| Certification File No.: | 01317 | Cert. Date: | 06/28/2023 |
| Linearity File No.: | 01318 | Lin. Date: | 06/28/2023 |
| Solution File No.: | 01331 | Soln. Date: | 09/13/2023 |
| Sequential File No.: | 01331 | File Date: | 09/13/2023 |
| | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDUH S3-0079 |
| Solution Control Lot: | 23010 | Expires: | 01/11/2025 |
| | | Bottle No.: | 0939 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 07:00D | | |
| Control 1 EC | 0.101% | 07:00D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.102% | 07:00D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:01D | | |
| Control 2 EC | 0.101% | 07:01D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 07:01D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:02D | | |
| Control 3 EC | 0.101% | 07:03D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.102% | 07:03D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:03D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: _____

PFC J M 408

Badge No.: 408

Date: 09/13/2023