

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 01316 Calib. Date: 06/28/2023 Calib. No.: 00053  
Certification File No.: 01317 Cert. Date: 06/28/2023 Cert. No.: 00040  
Linearity File No.: 01318 Lin. Date: 06/28/2023 Lin. No.: 00040  
Solution File No.: 01335 Soln. Date: 10/11/2023 Soln. No.: 00290  
Sequential File No.: 01335 File Date: 10/11/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079  
Control Solution %: 0.100% Expires: 01/11/2025  
Solution Control Lot: 23010 Bottle No.: 1005

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:37D		
Control 1 EC	0.101%	07:38D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	07:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:38D		
Control 2 EC	0.101%	07:39D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	07:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:40D		
Control 3 EC	0.101%	07:40D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	07:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: GASSMAN

First Name: JASON

MI:

Signature: \_\_\_\_\_

*PFC J M 408*

Badge No.: 408

Date: 10/11/2023