

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 01316 Calib. Date: 06/28/2023 Calib. No.: 00053
Certification File No.: 01317 Cert. Date: 06/28/2023 Cert. No.: 00040
Linearity File No.: 01318 Lin. Date: 06/28/2023 Lin. No.: 00040
Solution File No.: 01345 Soln. Date: 12/06/2023 Soln. No.: 00292
Sequential File No.: 01345 File Date: 12/06/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079
Control Solution %: 0.100% Expires: 04/12/2025
Solution Control Lot: 23110 Bottle No.: 0472

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 12:55S | | |
| Control 1 EC | 0.099% | 12:56S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 12:56S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:56S | | |
| Control 2 EC | 0.099% | 12:57S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 12:57S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:58S | | |
| Control 3 EC | 0.098% | 12:59S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 12:59S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:59S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 12/06/2023